

TECHNICAL DATA

Fluke 972A/972B/972ES Temperature Humidity Meter



Key features

- Standard and optional external probes are available for limited space
- Accuracy of ±0.5 °C and ± 3% RH, to meet industrial grade testing requirements
- 99 sets of stored data, easy to review (Fluke 972B only)
- Auto logging mode for up to 8 hours of automatic temperature and humidity monitoring (Fluke 972B only)
- Comfortable grip and ergonomic design
- Reinforced sensor protective cover for worry-free operation

Product overview: Fluke 972A/972B/972ES Temperature Humidity Meter

Measure-and-view, industrial design, necessity for inspection.

Temperature and humidity are two important items for comfort test of indoor air and environment. The Fluke 972 Series Temperature Humidity Meter can quickly and easily collect accurate humidity and temperature data. It features streamlined industrial design and is comfortable to hold. Simplified buttons require no complicated operations. The selected highprecision sensors ensure that reliable data is collected. The Fluke 972 Series is an indispensable environmental monitoring tool for facility maintenance and engineering technicians, HVAC service contractors, R&D and manufacturing engineers, and professionals involved in indoor air quality (IAQ) assessments.



Specifications: Fluke 972A/972B/972ES Temperature Humidity Meter

		972A	972B	972ES	
Detailed Specifications					
Sensor Type		Internal	Internal + External	External	
Temperature Measurement Range		-30.0 🛛 to 65.0 🕅	Internal sensor: -30.0 °C to 65.0 °C; External sensor: -40.0 °C to 99.9 °C	-40.0 °C to 99.9 °C	
Temperature Accuracy		±0.5 🛛 (-30.0 🖾 to 65.0 🖾)	Internal sensor: ±0.5 °C (-30.0 °C to 65.0 °C) External sensor: ±0.5 °C (-30.0 °C to 65.0 °C); ±0.7 °C (others)	±0.5 °C (-30.0 °C to 65.0 °C); ±0.7 °C (others)	
Temperature Resolution		0.1 °C			
Dew Point Temperature	Range	N/A	-40.0 °C to 99.9 °C	N/A	
	Resolution	N/A	0.1 °C	N/A	
Wet Bulb Temperature	Range	N/A	-20.0 °C to 60.0 °C	N/A	
	Resolution	N/A	0.1 °C	N/A	
Humidity	Range	0.0% RH to 99.9% RH			
	Accuracy	±3% RH (≥20.0% RH to ≤80.0% RH @ 25 ⊠) ±5% RH (≥1% RH to <20% RH, >80% RH to ≤95% RH @ 25⊠) Unspecified for others			
	Resolution	0.1% RH			
Data Storage		N/A	Yes. Up to 99 sets	N/A	
Display		2-line LCD display, with backlight		No	
Auto-off		Power off after 20 min without any operation		N/A	
Max/Min/Avg		Yes	Yes	N/A	
Auto Logging Mode		N/A	Yes	N/A	
General Specifications					
Temperature	Operating	-30 °C to 70 °C -40		-40 °C to 70 °C	
	Storage	-30 °C to 70 °C, <80% RH (without battery)		-40 °C to 70 °C, <80% RH	
Altitude	Operating	2000 m			
	Storage	12 000 m			
Relative Humidity	Operating	0% RH to 99% RH			
	Storage	180% RH			
Power Supply	Battery Type	2 x AAA Batteries –			
	Battery Life	1000 hrs for continuous operating, backlight off —			

2 Fluke Corporation Fluke 972A/972B/972ES Temperature Humidity Meter



Safety	IEC61010-1, Pollution Degree 2		
Dimensions (L x W x H)	211 × 58 × 40 mm	264 × 33 × 28 mm	
Weight	172 g (body)	127 g	
	IEC 61326-1: Portable, Electromagnetic Environment; IEC 61326-2-2; CISPR 11: Group 1; Class A		
Electromagnetic Environment (EMC)	Group 1: Equipment intentionally generates and/or uses conductive of energy, which is necessary for its internal operation.		
IP Rating	IEC 60529: IP30, non-operating		



Ordering information



Fluke 972A/972B/972ES

Fluke 972A/972B/972ES Temperature Humidity Meter

Fluke 972A Temperature Humidity Meter

Includes

• The Fluke 972A Temperature Humidity Meter

Fluke 972B Temperature Humidity Meter

Includes

- The Fluke 972B Temperature Humidity Meter
- Fluke 972ES external sensor

Fluke 972ES External Temperature/Humidity Sensor

Includes

• Fluke 972ES external sensor



Fluke. Keeping your world up and running.®

Fluke Corporation PO Box 9090, Everett, WA 98206 U.S.A.

For more information call:

In the U.S.A. (800) 443-5853 In Canada (800) 36-FLUKE From other countries +1 (425) 446-5500 www.fluke.com ©2024 Fluke Corporation. Specifications subject to change without notice. 09/2024

Modification of this document is not permitted without written permission from Fluke Corporation.